



## Ocean Optics USB4000 UV-VIS Spectrometer

### Ocean Optics Red Tide UV-VIS Spectrometer

#### (Order Code: SP-UV-VIS, SPRT-UV-VIS)

The UV -VIS spectrometer is a portable ultraviolet light and visible light spectrophotometer, combining a spectrometer and a light source/cuvette holder. The light source/cuvette holder may be detached and an optical fiber assembly attached to the spectrometer for emission spectrum experiments.

#### What is included with the UV-VIS Spectrometer?

- One spectrometer with light source/cuvette holder (Ocean Optics UV-VIS Spectrometer, Ocean Optics Red Tide UV-VIS Spectrometer)
- AC power supply
- One package of 15 UV clear plastic cuvettes and lids
- One USB cable
- User's Guide (this document)

#### Important Note

Logger *Pro* 3 (version 3.6 or newer) software is required. If you own a previous version of Logger *Pro* 3, you may upgrade the software free of charge. You need the LabQuest application version 1.1 or newer to use a LabQuest as a stand-alone device with a spectrometer. Visit <http://www.vernier.com/LabQuest/updates>.

#### Get Started

##### Using an Ocean Optics Spectrometer with Your Computer

1. Install Logger *Pro* 3 software (version 3.6 or newer) on your computer before using an Ocean Optics spectrometer.
2. Connect the AC power supply to the light source/cuvette holder of the UV-VIS spectrometer. **Important:** The AC power must be connected before connecting to the USB port.
3. Connect the spectrometer to a powered USB port on the computer or a powered USB hub.

The first time you connect a spectrometer, your computer may ask you a few questions. **Note:** Do not go online for device drivers. The device drivers were installed when you installed Logger *Pro* 3.

- a. Windows computers (Windows XP or Vista) Follow the New Hardware Wizard instructions to download the drivers automatically.
  - b. Macintosh computers (Mac OS 10.3 or newer) If it appears, follow the New Device instructions.
4. Start the Logger *Pro* 3 software on your computer.

#### Calibrate the Spectrometer for Measuring Absorbance or %T

1. Allow the UV-VIS spectrometer to warm up for a minimum of ten minutes.
2. Fill a cuvette about  $\frac{3}{4}$  full with distilled water, or the solvent being used in the experiment, to serve as the blank. After the spectrometer has warmed up, place the blank cuvette in the spectrometer. Align the cuvette so a clear side of the cuvette is facing the light source.
3. Choose Calibrate ► Spectrometer from the Experiment menu. Follow the instructions in the dialog box to complete the calibration. Click .

#### Collect Data

There are three general types of data collection that measure absorbance – absorbance vs. wavelength which produces a spectrum, absorbance vs. concentration for Beer's law experiments, and absorbance vs. time for kinetics experiment.

#### Measure the Absorbance Spectrum of an Aqueous Sample (Absorbance vs. Wavelength)

1. Fill a cuvette about  $\frac{3}{4}$  full of a sample of the solution to be tested. Place the sample in the spectrometer and click . Click  to end the data collection.
2. To store the spectrum data, choose Store Latest Run from the Experiment menu.

#### Conduct a Beer's Law Experiment (Absorbance vs. Concentration)

1. Measure an absorbance spectrum as described above.
2. Click on the Configure Spectrometer Data Collection button, .
3. Click Abs vs. Concentration. The wavelength of the maximum absorbance will be automatically selected ( $\lambda$  max). Click  to continue or click  and select a wavelength on the graph or in the list of wavelengths.
4. Place your first Beer's law standard solution in the spectrometer. Click  and then click . Enter the concentration of the sample and click . Repeat this step for the remaining standard samples. Click  to end the data collection.
5. Click linear fit, , to see the function for the standard solutions.
6. Place an unknown sample of solution in the spectrometer. Choose Interpolation Calculator from the Analyze menu. A helper box will appear, displaying the absorbance and concentration of the unknown. Click .

#### Conduct a Kinetics Experiment (Absorbance vs. Time)

1. Measure an absorbance spectrum as described above.
2. Click on the Configure Spectrometer Data Collection button, .
3. Click Abs vs. Time. The wavelength of maximum absorbance will be selected as before. Click  to continue or click  and select a wavelength on the graph or in the list of wavelengths.

- The default settings are 1 sample per second for 200 seconds. To change the data-collection parameters for your experiment, choose Data Collection from the Experiment menu and make the necessary changes. Click .
- Mix the reactants. Transfer ~2 mL of the reaction mixture to a cuvette and place the cuvette in the spectrometer. Click . Click  if you wish to end the data collection early.
- Click Curve Fit,  to calculate a function for your data.

## Using the Spectrometer to Measure Emission Spectra

You may use your spectrometer to measure the emission spectrum of a light source such as an LED or a gas discharge tube. To do so, you may want to purchase an optical fiber assembly (order codes: UV-VIS).

### Measure an Emission Spectrum

- Use a small screwdriver to loosen the two screws that connect the cuvette holder to the spectrometer (see Figure 1). Remove the cuvette holder and connect an optical fiber assembly to the spectrometer.



Figure 1

- Use a USB cable to connect the spectrometer to your computer.
- Start *Logger Pro 3*.
- Choose Change Units ► Spectrometer ► Intensity from the Experiment menu. Intensity is a relative measure. The spectrometer is not calibrated for measuring intensity.
- Aim the tip of the optical fiber cable at a light source. Click . Click  to end the data collection.

If the spectrum maxes out (flat and wide peaks), increase the distance between the light source and the tip of the optical fiber cable or reduce the sample time. To adjust the data collection parameters, choose Set Up Sensors ► Spectrometer from the Experiment menu. Set the Sample Time to a suitable value and decrease the Samples to Average to 1.

### Using the Stored Emissions Files in *Logger Pro 3*

*Logger Pro 3* contains a folder of emissions graphs from selected discharge tubes, including: argon, helium, hydrogen, mercury, oxygen, sodium, and xenon. You can

display and analyze these graphs without a spectrometer connected to your computer. Follow the steps below to view one of these graphs.

- Choose Open from the File menu.
- Open the Sample Data folder.
- In the Sample Data folder, open the Physics folder.
- In the Physics folder, open Gas Discharge Spectra. Open the desired file.

## Changing the Settings in *Logger Pro 3*

### Spectrometer Dialog Box

The Spectrometer dialog box lists all the settings for the device. To display this box choose Set Up Sensors ► Show All Interfaces from the Experiment menu.

For most experiments, the default settings work well.

There are four parameters listed in the dialog box.

- Sample Time: think shutter speed of a camera. *Logger Pro 3* automatically selects the proper sample time during calibration. **Note:** For emission studies, you may need to change the sample time manually.
- Wavelength Smoothing: the number of adjacent readings on either side of a given value that is used to calculate an average value.
- Samples to Average: the number of readings taken at a given wavelength to calculate an average reading.
- Wavelength Range: the range is determined by the type of spectrometer in use.

By clicking on the picture of the spectrometer in this dialog box, you will gain access to four options: calibrate, configure data collection, go to support web page, and units of measure. Click on an item to select it.

### Configure Spectrometer Data Collection Dialog Box

To display this box, click on the Configure Data Collection button, .

There are three regions in this box and four buttons at the bottom.

- Graph: The graph displays a full spectrum analysis of your sample. By default, the wavelength of greatest absorbance ( $\lambda$  max) will be marked with a box. You may select other wavelengths by clicking on the plot at the desired wavelength.
- Set Collection Mode: Three options for data collection are offered.
- Full Spectrum/Select Wavelength: This column of the available wavelengths becomes active when you select Abs vs. Concentration or Abs vs. Time. Check the box for each wavelength you wish to use in an experiment. When you select a wavelength from the list, a box appears on the graph.
- Use the  button to remove all of the wavelengths selected on the graph.

### Determining the Wavelength(s) to Use in an Experiment

When you conduct a Beer's law lab or a kinetics lab, it is common to select one wavelength at which to follow the experiment. However, in *Logger Pro 3* you may select as many wavelengths as you wish. There are two ways to select the wavelength or wavelengths.

### 1. Perform a Full Spectrum Analysis of the Solution to Be Tested

Measure the full spectrum of a sample of solution and then click Configure Spectrum Data Collection (). Select Abs vs. Concentration or Abs vs. Time. The wavelength of maximum absorbance ( $\lambda$  max) will be automatically selected.

### 2. Use a Sample of Solution to Determine the Peak Absorbance

This is a variation of the previous method. After calibrating the Spectrometer, place a sample of solution in the Spectrometer and then click Configure Spectrum Data Collection (). Select Abs vs. Time or Abs vs. Concentration. The wavelength of maximum absorbance ( $\lambda$  max) will be automatically selected. If you don't want to use the  $\lambda$  max, click  and select a wavelength on the graph or in the list of wavelengths.

### Selecting a Range of Wavelengths to Use in an Experiment

You may wish to measure the absorbance or %T of a sample over a group of wavelengths rather than a single wavelength. There are two ways to select a group of wavelengths from the Configure Spectrum Data Collection dialog box.

- Select the wavelengths one at a time by checking the boxes in the Select Wavelength column.
- Left click on the graph and drag your cursor across the region of wavelengths to analyze. Check the “Treat Contiguous Wavelengths as a Single Range” box.

### Measurement

You can set up the spectrometer to measure intensity, absorbance, or % transmittance. Choose Change Units ► Spectrometer from the Experiment menu. Click on the unit of choice from the list.

### Using an Ocean Optics Spectrometer with a LabQuest

1. Connect the AC power supply to the spectrometer.
2. Use the USB cable to connect the spectrometer to a LabQuest. Allow the spectrometer to warm up for fifteen minutes.
3. Turn on the LabQuest. The LabQuest app will launch automatically and the meter screen will be displayed.

### Calibrate the Spectrometer

1. Fill a cuvette about  $\frac{3}{4}$  full with distilled water and place it in the spectrometer. Align the cuvette so a clear side of the cuvette is facing the light source.
2. Choose Calibrate ► USB:Spectrometer from the Sensors menu. At the prompt, select Finish Calibration. After the message “Calibration Completed” appears, select .

### Measure the Absorbance Spectrum of an Aqueous Sample (Absorbance vs. Wavelength)

1. Fill a cuvette about  $\frac{3}{4}$  full of the solution to be tested and place it in the spectrometer.
2. Start the data collection. Tap the red Stop button to end the data collection.
3. **Note:** The wavelength of maximum absorbance ( $\lambda$  max) is automatically selected. This  $\lambda$  max will be used for any subsequent data collection, such as a Beer's Law experiment (abs vs. conc.) or a kinetics experiment (abs. vs. time). If you wish to choose another wavelength, you can tap on the graph to select a new wavelength or you can use the arrow keys on the keypad to move the cursor to a new wavelength.

### Conduct a Beer's Law Experiment (Absorbance vs. Concentration)

1. Measure an absorbance spectrum as described above. On the Meter screen, tap Mode. Change the mode to Events with Entry.
2. Enter the Name (Concentration) and Units (mol/L). Select OK.
3. A message will appear warning you to either save or discard the full spectrum run. Make your choice and proceed with the data collection.
4. Place your first Beer's law standard solution in the spectrometer. Start data collection. After the absorbance reading stabilizes, tap Keep. Enter the concentration of the solution and select OK.
5. Place your second standard sample in the spectrometer. After the absorbance readings stabilize, tap Keep. Enter the concentration of the second sample and select OK.
6. Repeat Step 5 for the remaining standard samples. After you have tested the final standard, tap the red Stop button to end the data collection.
7. To calculate a best fit line equation for your standards, choose Curve Fit from the Analyze menu. Select Linear for the Fit Equation, and then select OK. The graph screen will appear again with the linear regression equation displayed.
8. Place a cuvette containing an unknown sample of solution in the spectrometer. Tap the Meter tab and write down the displayed absorbance value. Tap the graph tab and trace the linear regression equation to determine the concentration of the unknown.

### Conduct a Kinetics Experiment (Absorbance vs. Time)

1. Measure an absorbance spectrum as described above. On the Meter screen, tap Mode. Change the data-collection mode to Time Based.
2. You can change the rate, interval, and/or length of time of data collection, if desired. Select OK when you are ready to proceed.
3. A message will appear warning you to either save or discard the full spectrum run. Make your choice and proceed with data collection.

- Mix the reactants, transfer ~2 mL of the reaction mixture to a cuvette and place the cuvette in the spectrometer. Start the data collection. You may tap the red Stop button to end the data collection early.
- To calculate a function for your data, choose Curve Fit from the Analyze menu. Select the Fit Equation, and then select OK. The graph screen will appear again.

### Measure an Emission Spectrum

- Use a small screwdriver to loosen the two screws that connect the cuvette holder to the Spectrometer (see Figure 1). Remove the cuvette holder and connect an optical fiber assembly to the spectrometer.
- Turn on the LabQuest. The LabQuest App will launch automatically and the meter screen will be displayed.
- On the meter screen, tap Change Units ► USB:Spectrometer ► Intensity from the Sensors menu. The spectrometer measures intensity in relative units.
- Aim the tip of the optical fiber cable at a light source. Start the data collection. Tap the red Stop button to end the data collection.

If the spectrum maxes out (flat and wide peaks), increase the distance between the light source and the tip of the optical fiber cable or reduce the sample time. To adjust the data collection parameters, tap Sensors and choose Data Collection. Set the Sample Time to a suitable value and decrease the Samples to Average to 1.

### Determining the Wavelength to Use in an Experiment

After you collect a full absorbance spectrum of a sample, LabQuest will identify the wavelength of maximum absorbance ( $\lambda_{max}$ ). If you wish to select a different wavelength, tap on the full spectrum graph or use the arrow keys on the keypad to identify the wavelength of choice. Another way to change the wavelength is to navigate to the meter screen, tap on the meter itself, and select Change Wavelength. Enter the wavelength of your choice and select OK. If the wavelength you type in is not measured by the UV-VIS unit, LabQuest will automatically choose the wavelength closest to your choice.

### Measurement

You can set up the spectrometer to measure intensity, absorbance, or % transmittance. On the Meter screen, choose Change Units from the Sensors menu. Click on the unit of choice from the list.

### Sample Experiments

There are several experiments available for use with the spectrometer. You may download the labs from our web site ([www.vernier.com/spectroscopy](http://www.vernier.com/spectroscopy)).

### Specifications

#### Ocean Optics UV-VIS Spectrometer (order code: SP-UV-VIS)


Dimensions: 14.2 cm × 8.7 cm × 3 cm (includes cuvette holder/light source)  
 Power: AC adapter (included)  
 Light Source: integrated deuterium tungsten halogen light source  
 Wavelength Range: 200 nm–850 nm  
 Pixel Resolution: 0.2 nm

#### Ocean Optics Red Tide UV-VIS Spectrometer (order code: SPRT-UV-VIS)

Dimensions: 14.2 cm × 8.7 cm × 3 cm (includes cuvette holder/light source)  
 Power: AC adapter (included)  
 Light Source: integrated deuterium tungsten halogen light source  
 Wavelength Range: 200 nm–850 nm  
 Pixel Resolution: 1 nm

### Warranty

Vernier warrants this product to be free from defects in materials and workmanship for a period of one year from the date of shipment to the customer. This warranty does not cover damage to the product caused by abuse or improper use. This product is manufactured by Ocean Optics, Inc.



#### Vernier Software & Technology

13979 S.W. Millikan Way • Beaverton, OR 97005-2886  
 Toll Free (888) 837-6437 • (503) 277-2299 • FAX (503) 277-2440  
[info@vernier.com](mailto:info@vernier.com) • [www.vernier.com](http://www.vernier.com)

Rev. 8/8/08

Logger Pro, Vernier LabPro, Go! Link and other marks shown are our registered trademarks in the United States. All other marks not owned by us that appear herein are the property of their respective owners, who may or may not be affiliated with, connected to, or sponsored by us.



Printed on recycled paper.